



AMERICAN BOTTOMS REGIONAL WASTEWATER TREATMENT FACILITY

MEMORANDUM

TO: All Potential Bidders

FROM: Charlie Loftus, Laboratory Supervisor

DATE: June 13, 2017

SUBJECT: Addendum to Bid AB 32-18 – Two Corrections

The first correction is found in section 4.a. in which the term "simultaneous" needs to be removed and sentence clarified. The corrected specification in 4.a. states as follows:

"The instrument detector must be a solid state device capable of testing a sample using radial and axial wavelengths in one analytical run."

The second correction is the operating system requirement in section 10.b. which states only Windows 10 Professional. This is not a necessary requirement, although Windows 10 Professional is preferable, the corrected specification now reads as follows for section 10.b.:

"The operating system must be Windows 7 or newer."